



TTs Serisi Tork Torna Vidaları - [0.3 - 6.0 N.m]

Torque Screwdrivers - [0.3 - 6.0 N.m]



- » Torklama hassasiyeti $\pm\% 6$
 - » Tork aralığı **0.3 - 6.0 N.m / 30 - 600 cN.m**
 - » Tork aralığı için **2.5 - 56 Lbf.in**
 - » ISO 6789-1:2017 standardına uygun
 - » Her birim için 3 farklı tork aralığı
 - » 1/4" içten altıgen sürücü ile sevk edilir
 - » Tork ayarı kabzadaki mandal ile yapılır, kilitlenir
 - » Ayarlanan tork değerinin üzerine çıkmaz
 - » Fabrika kalibrasyon belgesi ile sevk edilir
 - » Türkak akreditasyonlu kalibrasyon opsiyoneldir
 - » Menşei İngiltere, üretici Norbar, CE sertifikalı
 - » Servis ve kalibrasyon Surkon tarafından karşılanır
-
- » Torque accuracy $\pm 6\%$
 - » Torque range 0.3 - 6.0 N.m / 30 - 600 cN.m
 - » Torque range 2.5 - 56 Lbf.in
 - » Accuracy meets requirements of ISO 6789-1:2017
 - » Single scale either N.m or lbf.in selectable
 - » Adjustable or production type "P" available
 - » Three different torque ranges for each model
 - » Supplied with 1/4" female hexagon bit holder
 - » Does not exceed the selected torque value
 - » Supplied with house calibration certificate
 - » Manufactured in UK, Brand Norbar, CE certified

N.m Göstergeli Ayarlanabilir Tork Torna Vidaları

N.m Scale Adjustable Torque Screwdrivers

Parça Part	Model Model	Tork Aralığı Torque Range		Ölçek Graduation	Uzunluk Length	Ağırlık Weight
#	#	N.m	cN.m	N.m	mm	kg
13850	TTs1.5	0.3-1.5	30-150	0.05	155	0.2
13851	TTs3.0	0.6-3.0	60-300	0.10	155	0.2
13852	TTs6.0	1.2-6.0	120-600	0.20	155	0.2

İmalat için Sabit Ayarlı "P" Tip Tork Torna Vidaları

Production "P" Type Screwdrivers

Parça Part	Model Model	Tork Aralığı Torque Range		Uzunluk Length	Ağırlık Weight
#	#	N.m	lbf.in	mm	kg
13856	TTs1.5	0.3-1.5	2.5-13	155	0.2
13857	TTs3.0	0.6-3.0	5.0-26	155	0.2
13858	TTs6.0	1.2-6.0	10-53	155	0.2



Lbf.in Göstergeli Ayarlanabilir Tork Torna Vidaları

Lbf.in Scale Adjustable Torque Screwdrivers

Parça Part	Model Model	Tork Aralığı Torque Range		Ölçek Increment	Uzunluk Length	Ağırlık Weight
#	#	N.m	lbf.in	lbf.in	mm	kg
13853	TTs13 in.lb	*	2.5-13	1.0	155	0.2
13854	TTs26 in.lb	*	5-26	1.0	155	0.2
13855	TTs53 in.lb	*	10-53	2.5	155	0.2

Tork Torna Vidası ve Bits Uç Setleri

Torque Screwdriver & Bits Set



- » **13700 - TTS 0.3-1.5 KIT**
Tork Torna Vidası Seti 12 adet bits ucu
Torque Screwdriver kit with 12 bits
- » **13701 - TTs 0.6-3.0 KIT**
Tork Torna Vidası Seti 12 adet bits ucu
Torque Screwdriver kit with 12 bits
- » **13702 - TTs 1.2-6.0 KIT**
Tork Torna Vidası Seti 12 adet bits ucu
Torque Screwdriver kit with 12 bits

Set içerisindeki Uçlar/Bits included in the kit

- » T10 Torx
- » T15 Torx
- » T20 Torx
- » PH.1 Phillips Yıldız
- » PH.2 Phillips Yıldız
- » 3mm Düz (Flat)
- » 1/4" Erkek Kare (Male Sq.)
- » 2mm Erkek Altıgen (Male Hex.)
- » 2.5mm Erkek Altıgen (Male Hex.)
- » 3mm Erkek Altıgen (Male Hex.)
- » 4mm Erkek Altıgen (Male Hex.)
- » 5mm Erkek Altıgen (Male Hex.)



Norbar Torque Tools Ltd	
Certificate of Calibration	
Model: 13858	Serial No: 3012300007
Maximum Capacity: 60.0	Units: N.m
Inspector: BK	Ambient Temperature: 20°C
Actual Readings	
Torque	Min Max
10.0	9.82 10.42 10.10 10.00 10.10 10.00 10.00
20.0	20.85 21.25 20.81 20.48 20.46 20.48 20.47
50.0	48.08 52.08 50.18 50.17 50.16 50.14 50.14
The certificate and the test equipment used for this calibration comply with the requirements of ISO/IEC 17025	
The uncertainty of the test equipment at 20°C is $\pm 1\%$ k=2	
This test equipment used in the performance of the above calibration has international ISO/IEC 17025 accreditation for the following performance evaluation by UKAS accreditation	
UKAS Laboratory No: 1056	Date of Calibration: 8 Nov 2015
Tester Name: (Date) TD	Quality Manager: J.Boone
Serial No: 10000	
Cal No: 00000	